

# झारखण्ड केन्द्रीय विश्वविद्यालय Central University of Jharkhand

Brambe, Ranchi, Jharkhand, 835205

Dated: 24-12-2014

CUJ/C0E-GEET/NIT/2014/1

Sealed quotations are invited for the articles enclosed in the separate sheet subject to the termsand conditions given below. The quotation should be addressed to Prof. S.K. Samdarshi, (P.I. of the Project), Centre for Energy Engineering, Central University of Jharkhand, Ratu-Lohardaga Road, Brambe, Ranchi, 835205 and should reach on or before 27/02/2015. Quotations will be opened on the next working day.

#### TERMS AND CONDITIONS

- 1. Quotations without any erasures and overwriting must be submitted in sealed cover addressed to the, Prof. S.K.Samdarshi, Centre for Energy Engineering, Central University of Jharkhand, Ratu-Lohardaga road, Brambe, Ranchi- 835 205 super- scribed Tender/ Enquiry No. and thedue date failing which, quotation may be ignored. Tender/ quotation should be sent throughPost/ Courier/ By Hand.
- 2. The rate quoted should be inclusive of all packing, forwarding, commissioning & installation, sales tax, freight and insurance charges and should remain valid for our acceptance forminimum period of six (6) months from the due date of opening of the quotations. Thequotation should be CUJ consignee's site basis i.e. Central University of Jharkhand, RatuLohardagaRoad, Brambe, Ranchi.
- 3. Manufacturer's name of company of origin of materials offered must be clearly specified. Complete details of illustrated literatures/ or drawings, in original (not photocopies), if anymust accompany all quotations.
- 4. The University will not entertain requests for revision in prices once quoted for whateverreason after the tenders are opened during the period of contract.
- 5. All goods must be delivered at our University at Ratu- Lohardaga Road, Brambe, Ranchi for inspection by our inspecting authority.
- 6. Full payment will be made within 30 days of the receipt and acceptance after inspection of supplies at destination. The quotations containing different payment terms are liable to be ignored.
- 7. All goods must be delivered at our university at Ratu- Lohardaga road, Brambe, Ranchi(Stores & Purchase Section) after confirmed order. The University will provide DSIRcertificate, Educational certificate if necessary. The University will not be responsible for ssuing any Road Permit.
- 8. Only Manufacturer/ Authorized dealer/ firms of repute dealing in the items listed in the

- quotation having experience in sale and repair/ maintenance, need apply.
- 9. Only latest models/mentioned models of items need to be quoted.
- 10. It should be mentioned specifically whether price quoted includes all taxes and duties. Salestax and/ or other duties legally liable and intended to be claimed should be distinctly shown in the tender.
- 11. VAT/CST Registration Number and its validity should be indicated. Documentary evidences be supplied/ attached to the tender documents.
- 12. All rates shall be indicated both in words and figures. Where there is difference between the rates quoted in words and figures, rate quoted in words will prevail.
- 13. The supplier shall make delivery of the items within the stipulated period from the placement of the purchase order. The purchase order would be placed after assessing the requirementand hence, actual quantity may increase/ decrease
- 14. The Tender Document for items will be Two- Bid system consisting of Technical Bid and Price Bid. The Tender Document will be submitted item-wise in two separate sealed coversclearly mentioning on the envelope the details of items for which Bid is submitted.
- 15. The Bidder must enclose point wise compliance of the technical specifications of each item.
- 16. The price of Tender Document is Rs. 1000/- only (One Thousand rupees only) (nonrefundable). The Tender Document price may be attached to the Technical Bid in the form of Bank Draft in favour of "Central University of Jharkhand" payable at Ranchi.
- 17. The Tender Documents comprising of Price- Bid, Technical Bid and Earnest Money Deposit (refundable)mentioned against each item in the form of Bank Draftin favour of "Central University of Jharkhand" payable at Ranchi kept in the separateenvelope super- scribed with the name of the same must be deposited. The Tender Document must be enclosedwith Earnest Money Deposit otherwise the tender document will be rejected. All thedocuments must be enclosed in a bigger size envelope super- scribed with the tender numberand item name.
- 18. Only those Manufacturers liable to participate having ISO 9001 and ISO 14001 certification.
- 19. All legal disputes shall be under the jurisdiction of Jharkhand High court, Ranchi.
- 20. The University reserves the right to accept or reject any Bid, without assigning any reason thereof. No correspondence in this regard will be entertained.
- 21. All the participating firms should attach supporting documents in favour of:
  - a) Have at least 5 years of experience in the field.
  - b) Have an annual turnover of at least Rs. 50 Lakh per annum for each of the last 3 years (i.e., 2011-12, 2012-13& 2013-14) (Agency should produce a certificate from their Audit Firmin respect of turnover for the mentioned period)
  - c) Be registered with Directorate of Industries or Small Scale Industries or any otherappropriate authorities.
  - d) Have valid TIN Number for registration under VAT/CST

- e) Have PAN for Income Tax. Supporting documents for Income Tax return for the last Twoyears.
- f) Documents as proof in respect of Technical bid along with supporting documents.
- g) Profile and Track Record of the agency (3 years)
- h) Copy of signed Balance Sheets of last three years.
- i) Detailed terms & conditions regarding performance/ warranty/ bank guarantee/ Inspection/LD/ penalty for delayed supplied will be elaborated in purchase order.For proprietary items/ single manufacturers or Government manufacturers, propercertification in original is to be attached along with the Technical Bid.

# **List of Instruments**

Tube Load: 300 − 600 W Tube Voltage: 20 -40 kV Tube current: 10-30 mA Radiation enclosure: Included Source: Cu Scanning Method: □ - 2□ Sample holder: Different sample holders Filter: Kβ elimination and other aprropriate Filter Detector: Appropriate detector Cooling arrangement: Proper cooling arrangement; Other: UPS for XRD with 1h back up; Computer/printer with UPS for 30 min backup; Software: Standard Softwares	(in Rs.) 80000.00
Tube Voltage: 20 -40 kV Tube current: 10-30 mA Radiation enclosure: Included Source: Cu Scanning Method: □ - 2□ Sample holder: Different sample holders Filter: Kß elimination and other aprropriate Filter Detector: Appropriate detector Cooling arrangement: Proper cooling arrangement; Other: UPS for XRD with 1h back up; Computer/printer with UPS for 30 min backup; Software: Standard Softwares  2 UV-VIS-NIR Spectrophotometer with Diffuse Reflectance Attachment(UV-VIS-NIR) NIR-DRS) Wavelength Range: 190-3300 nm Beam: Double Beam Spectral Bandwidth: both for UV-Vis and NIR in ~ fraction of nm range Cell: 1.0 ml and 3.0 ml quartz (UV/Vis/NIR) Sample holder: Powder and Solid sample holder Software: Suitable for the instrument Attachment: Suitable integrating sphere for DRS	
Tube current: 10-30 mA Radiation enclosure: Included Source: Cu Scanning Method: □ - 2□ Sample holder:Different sample holders Filter:Kß elimination and other aprropriate Filter Detector:Appropriate detector Cooling arrangement: Proper cooling arrangement; Other:UPS for XRD with 1h back up; Computer/printer with UPS for 30 min backup; Software:Standard Softwares  2 UV-VIS-NIR Spectrophotometer with Diffuse Reflectance Attachment(UV-VIS-NIR-DRS) Wavelength Range: 190-3300 nm Beam: Double Beam Spectral Bandwidth: both for UV-Vis and NIR in ~ fraction of nm range Cell:1.0 ml and 3.0 ml quartz (UV/Vis/NIR) Sample holder: Powder and Solid sample holder Software: Suitable for the instrument Attachment: Suitable integrating sphere for DRS	40000.00
Source : Cu Scanning Method : □ - 2□ Sample holder:Different sample holders Filter:Kß elimination and other aprropriate Filter Detector:Appropriate detector Cooling arrangement: Proper cooling arrangement; Other:UPS for XRD with 1h back up; Computer/printer with UPS for 30 min backup; Software:Standard Softwares  2 UV-VIS-NIR Spectrophotometer with Diffuse Reflectance Attachment(UV-VIS-NIR-DRS) Wavelength Range: 190-3300 nm Beam: Double Beam Spectral Bandwidth: both for UV-Vis and NIR in ~ fraction of nm range Cell:1.0 ml and 3.0 ml quartz (UV/Vis/NIR) Sample holder: Powder and Solid sample holder Software: Suitable for the instrument Attachment: Suitable integrating sphere for DRS	40000.00
Scanning Method: □ - 2□ Sample holder:Different sample holders Filter:Kß elimination and other aprropriate Filter Detector:Appropriate detector Cooling arrangement: Proper cooling arrangement; Other:UPS for XRD with 1h back up; Computer/printer with UPS for 30 min backup; Software:Standard Softwares  2 UV-VIS-NIR Spectrophotometer with Diffuse Reflectance Attachment(UV-VIS-NIR-DRS) Wavelength Range: 190-3300 nm Spectral Bandwidth: both for UV-Vis and NIR in ~ fraction of nm range Cell:1.0 ml and 3.0 ml quartz (UV/Vis/NIR) Sample holder: Powder and Solid sample holder Software: Suitable for the instrument Attachment: Suitable integrating sphere for DRS	40000.00
Sample holder: Different sample holders Filter: Kß elimination and other aprropriate Filter Detector: Appropriate detector Cooling arrangement: Proper cooling arrangement; Other: UPS for XRD with 1h back up; Computer/printer with UPS for 30 min backup; Software: Standard Softwares  2 UV-VIS-NIR Spectrophotometer with Diffuse Reflectance Attachment(UV-VIS-NIR-DRS) Spectral Bandwidth: both for UV-Vis and NIR in ~ fraction of nm range Cell: 1.0 ml and 3.0 ml quartz (UV/Vis/NIR) Sample holder: Powder and Solid sample holder Software: Suitable for the instrument Attachment: Suitable integrating sphere for DRS	40000.00
Filter:Kß elimination and other aprropriate Filter Detector:Appropriate detector Cooling arrangement: Proper cooling arrangement; Other:UPS for XRD with 1h back up; Computer/printer with UPS for 30 min backup; Software:Standard Softwares  2 UV-VIS-NIR Spectrophotometer with Diffuse Reflectance Attachment(UV-VIS-NIR-DRS) Wavelength Range: 190-3300 nm Beam: Double Beam Spectral Bandwidth: both for UV-Vis and NIR in ~ fraction of nm range Cell:1.0 ml and 3.0 ml quartz (UV/Vis/NIR) Sample holder: Powder and Solid sample holder Software: Suitable for the instrument Attachment: Suitable integrating sphere for DRS	40000.00
Detector: Appropriate detector Cooling arrangement: Proper cooling arrangement; Other: UPS for XRD with 1h back up; Computer/printer with UPS for 30 min backup; Software: Standard Softwares  2 UV-VIS-NIR Spectrophotometer with Diffuse Reflectance Attachment(UV-VIS-NIR-DRS) Wavelength Range: 190-3300 nm Beam: Double Beam Spectral Bandwidth: both for UV-Vis and NIR in ~ fraction of nm range Cell: 1.0 ml and 3.0 ml quartz (UV/Vis/NIR) Sample holder: Powder and Solid sample holder Software: Suitable for the instrument Attachment: Suitable integrating sphere for DRS	40000.00
Cooling arrangement: Proper cooling arrangement; Other:UPS for XRD with 1h back up; Computer/printer with UPS for 30 min backup; Software:Standard Softwares  2 UV-VIS-NIR Spectrophotometer with Diffuse Reflectance Attachment(UV-VIS-NIR-DRS) Wavelength Range: 190-3300 nm Beam: Double Beam Spectral Bandwidth: both for UV-Vis and NIR in ~ fraction of nm range Cell:1.0 ml and 3.0 ml quartz (UV/Vis/NIR) Sample holder: Powder and Solid sample holder Software: Suitable for the instrument Attachment: Suitable integrating sphere for DRS	40000.00
Other: UPS for XRD with 1h back up; Computer/printer with UPS for 30 min backup; Software: Standard Softwares  2 UV-VIS-NIR Spectrophotometer with Diffuse Reflectance Attachment(UV-VIS-NIR-DRS) Sample holder: Powder and Solid sample holder Software: Suitable integrating sphere for DRS	40000.00
with UPS for 30 min backup; Software:Standard Softwares  2 UV-VIS-NIR Spectrophotometer with Diffuse Reflectance Attachment(UV-VIS-NIR-DRS)  Spectral Bandwidth: both for UV-Vis and NIR in ~ fraction of nm range Cell:1.0 ml and 3.0 ml quartz (UV/Vis/NIR) Sample holder: Powder and Solid sample holder Software: Suitable for the instrument Attachment: Suitable integrating sphere for DRS	40000.00
Software:Standard Softwares  UV-VIS-NIR Spectrophotometer with Diffuse Reflectance Attachment(UV-VIS-NIR-DRS)  NIR-DRS)  Software:Standard Softwares  Wavelength Range: 190-3300 nm Spectral Bandwidth: both for UV-Vis and NIR in ~ fraction of nm range Cell:1.0 ml and 3.0 ml quartz (UV/Vis/NIR) Sample holder: Powder and Solid sample holder Software: Suitable for the instrument Attachment: Suitable integrating sphere for DRS	40000.00
2 UV-VIS-NIR Spectrophotometer with Diffuse Reflectance Attachment(UV-VIS-NIR-DRS) Wavelength Range: 190-3300 nm Beam: Double Beam Spectral Bandwidth: both for UV-Vis and NIR in ~ fraction of nm range Cell:1.0 ml and 3.0 ml quartz (UV/Vis/NIR) Sample holder: Powder and Solid sample holder Software: Suitable for the instrument Attachment: Suitable integrating sphere for DRS	40000.00
Spectrophotometer with Diffuse Reflectance Attachment(UV-VIS- NIR-DRS)  Beam: Double Beam Spectral Bandwidth: both for UV-Vis and NIR in ~ fraction of nm range Cell:1.0 ml and 3.0 ml quartz (UV/Vis/NIR) Sample holder: Powder and Solid sample holder Software: Suitable for the instrument Attachment: Suitable integrating sphere for DRS	40000
Diffuse Reflectance Attachment(UV-VIS-NIR-DRS)  Spectral Bandwidth: both for UV-Vis and NIR in ~ fraction of nm range Cell:1.0 ml and 3.0 ml quartz (UV/Vis/NIR) Sample holder: Powder and Solid sample holder Software: Suitable for the instrument Attachment: Suitable integrating sphere for DRS	
Attachment(UV-VIS-NIR-DRS)  fraction of nm range Cell:1.0 ml and 3.0 ml quartz (UV/Vis/NIR) Sample holder: Powder and Solid sample holder Software: Suitable for the instrument Attachment: Suitable integrating sphere for DRS	
NIR-DRS)  Cell: 1.0 ml and 3.0 ml quartz (UV/Vis/NIR)  Sample holder: Powder and Solid sample holder  Software: Suitable for the instrument  Attachment: Suitable integrating sphere for DRS	
Software: Suitable for the instrument Attachment: Suitable integrating sphere for DRS	
Attachment: Suitable integrating sphere for DRS	
attachment	
3 Fluorescence PC controlled Fluorescence spectrophotometer for 3	20000 00
3   Fluorescence   PC controlled Fluorescence spectrophotometer for   3   Research purpose with ozone free xenon lamp and power	30000.00
supply.	
Entrance: Continuously adjustable entrance and exit slits	
(no stepwise) operated under computer control.	
Operational specification: Capable of automatic	
acquisition of corrected emission and excitation spectra,	
synchronous luminescence spectra, and kinetic studies.	
Upgradability:Further possible upgradation for TCSPC	
(Time Correlated Single Photon Counting) technique for	
lifetime measurement. Future possible attachment of	
polarizer system.	
The instrument will have the following detailed	
technical specification: Detection: Photon counting technique	
Optics: All reflective optics for focusing at all	
wavelengths	
Excitation source: Ozone free xenon arc lamp	
Excitation and Emission Monochromator: Czerny-	
Turner design	
Excitation range: 200 - 900 nm	
Emission range: 200 - 900 nm	
<b>Band pass</b> : 0 to 30 nm, continuously tuneable through	
software	
Wavelength accuracy: ±0.6 nm or better	
Scan speed: variable with maximum 80 nm/sec	
Integration time: 1 ms to 150 sec Emission detector: Photomultiplier Tube, 200 - 850 nm	
Reference detector: Silicon photodiode for excitation	
correction	
Sensitivity: >5,000:1 (FSD value) for water Raman signal	
(Ex: 350 nm, <b>Bandpass:</b> 5 nm, Integration time: 1 sec.)	

		Additional accessories need to be supplied  (a) Quartz Cuvette open top with lid, 10mm path length, volume 3 ml (2 nos).  (b) Set of out off filter (5 nos) with filter holder.	
4.	Atomic Force Microscope (AFM)	(b) Set of cut-off filter (5 nos.) with filter holder  Scanning Probe Microscope(SPM) modes included:  In Air:  Contact AFM Lateral Force Mode (LFM) Resonant Mode AFM (semicontact + noncontact AFM) Phase Imaging Force Modulation Microscopy Adhesion Force Imaging Scanning Tunneling Microscopy AFM/STM spectroscopy Magnetic Force Mode and Electrostatic Force Mode (MFM/EFM) AFM Nano-Lithography and Nanomanipulation (Force and Electro-oxidative) Spreading Resistance Imaging (SRI) Scanning Capacitance Imaging (SCI) Scanning Kelvin Probe Microscopy (SKM) Piezo Response Force Microscopy AFM spectroscopies STM spectroscopies Single XYZ peizo scanner: Scanners must be single peizo tube in comparison to decoupled or flexure based scanners for highest resolution, low noise & high scanning speed. No decoupled or flexure stage scanners will be considered. Scanning type - XYZ by sample without any motion of the tip in XYZ Scanning range: 100x100x10um. Must be able to operate in both open and closed loop. The same scanner must have capability to work as small area scanner < 5x 5 um for critical resolution. Scanner should low voltage mode with XYZ range of 3x3x2 um. All modes should be performed by closed loop controlled scanner	80000.00
		SPM Controller:  Digital signals control Possibility of using several feed backs simultaneously Up to 8 signals should be detected during single pass DSP processor 320 MHz floating-point	

- USB interface with PC
- Integrated closed loop control
- +/- 10 V module for Bias Voltage control and lithography operations
- Force constant probe calibration integrated
- Three 16 bit controllers for each channel (X,Y,Z).
- Controller must provide lock-in capable of 360 degree quantitative cantilever phase detection, 0 Hz to 5 MHz.

### Scanner Closed-loop control and noise level:

- Non-linearity of XY: 0.1%
- Non-linearity of Z: 1%;
- Noise level of XY: 0.3 nm (Closed-loop on)0.02 nm (Closed loop off)
- Noise level of Z: 0.03 nm (Closed-loop on) 0.02 nm (Closed loop off)

#### Easy to use software and automation:

- Automated software driven control of measurement modes
- Automated tip-to-sample approach
- Fully Automated AFM and STM measuring heads exchange
- Automated laser/photodiode alignment of optical feedback geometry
- Motorized software driven sample positioning in XY, at least 5x5 mm range
- Motorized focus and continuous zoom of the optical viewing system with resolution 3um or better
- Must have universal cantilever holder.

#### **Computer & Workstation:**

A compatible desktop computer with monitor (23") with latest Windows operating system with high speed processor and Video capture capability for onscreen.

## Accessories/Manuals/Warranty:

- At least 50 probes for above mentioned AFM applications/modes
- Set of calibration standards for calibration piezo scanners
- Complete toolkit enough to operate with device
- English version of operating and service manuals
- System should have at least 1 years warranty
- Software update should be free of charge for life time device use.

- Company should provide local service support
The following options should be available for upgrade: (OPTIONS)  Heating stage
<ul> <li>Possibility to operate with up to 150 °C temperature range;</li> <li>Temperature stability - 0.05 °C.</li> </ul>
Hardware Nanoindentation  - Hardness measurement in the range 1-80 GPa;  - Young Modulus measurement in the range 1-1000 GPa;
- Measurements must be performed by Berkovich type diamond probe with optical force sensor

To

The Registrar

CUJ, Ranchi

Sub: NIT for some instruments to be purchased for CoE-GEET

Sir

With reference to the above we would like to request you to kindly arrange to invite tenders for the following instruments

- i) Bench Top X-ray Diffractometer (XRD)
- ii) UV-VIS-NIR Spectrophotometer with Diffuse Reflectance Attachment (UV-VIS-NIR-DRS)
- iii) Photoluminescence Spectrophotometer(PL)

The notification and detailed specification for each one of them is provided in the attached document. It is requested to get the same uploaded on the CUJ website with the notice, which are attached herewith.

Thanks

S K Samdarshi Coordinator and PI